



شركة الرومي للتجارة العامة ذ م م

AL ROUMI

GENERAL TRADING COMPANY L L C

www.alroumiae.com

AL ROUMI GENERAL TRADING Co. LLC.

H2S SAFETY DIVISION

P.O. BOX 45333, Abu Dhabi, UAE

Tel # (02) 6729901 : Fax # (02) 6729771

MULTIGAS DETECTOR CALIBRATION CERTIFICATE

Certificate No-	AR-MD-6/2026-1		
Company	NCMT		
Location	AL ROUMI BASE		
Instrument	Multigas Detector		
Model BW Technologies	BW Micro Clip - MA224-005489		
Sensor Details	TOXIC 1	Sensor Details	TOXIC 2
Sensor Fitted	H2S	Sensor Fitted	CO
Alarm LOW	10 PPM	Alarm LOW	35 PPM
Alarm HIGH	15 PPM	Alarm HIGH	200 PPM
TWA	10 PPM	TWA	35 PPM
STEL	15 PPM	STEL	50 PPM
STATUS	OK	STATUS	OK
Sensor Details	FLAMMABLE 1	Calibration Gas Details	
Sensor Fitted	LEL	Methane	2.5%Vol
Alarm LOW	10.0% LEL	Oxygen	18.0%Vol
Alarm HIGH	20.0% LEL	H2S	25 ppm
STATUS	OK	CO	100 ppm
-	-	Batch Number	11926
-	-	Expiry Date	Sep-26
Sensor Details	OXYGEN	Detecting Range	
Sensor Fitted	O2	Oxygen (O2)	0-30%
Alarm LOW	19.50%	Combustible Gas (LEL)	0-100%
Alarm HIGH	23.50%	Carbon Monoxide (CO)	0-1000 PPM
STATUS	OK	Hydrogen Sulphide (H2S)	0-200 PPM
Sensor Details	SULFUR DIOXIDE	Calibration Gas Details	
Sensor Fitted	N/A	Sulfur Dioxide	N/A
Alarm LOW	N/A	Balance	N/A
Alarm HIGH	N/A	Batch Number	N/A
STATUS	N/A	Expiry Date	N/A
Calibrated on:	18-Jun-26	Calibration Due on	14-Dec-26
Calibrated by	BINEESH B	Noted by	SANDEEP
Position	H2S ENGINEER	Position	QHSE SUPERVISOR
Signature		Signature	