PanReac AppliChem ITW Reagents

ANALYSIS CERTIFICATE

CODE: **273618** **BATCH:** 0002196890

PRODUCT: **Buffer Solution pH 10.00 (20°C) (blue colour)**

ISSUE DATE: 14/09/2022 **RETEST DATE:** 09/2028

SPECIFICATIONS	GUARANTEE VALUE	ACTUAL VALUE
pH value at 20°C	9,95 - 10,05	10,00
Uncertainty	± 0,015	

Determination

The determination is carried out with a glass combined electrode, potentiometrically, by linear interpolation on a multiple point calibration, obtained with SRM of NIST:

SRM 189c - Potassium Tetraoxalate
SRM 188 - Potassium Hydrogen Tartrate
SRM 185h - Potassium Hydrogen Phtalate
SRM 186 Ig - Potassium di-Hydrogen Phosphate
SRM 186IIg - di-Sodium Hydrogen Phosphate
SRM 187e - Sodium Tetraborate decahydrate
Calcium Hydroxide according to DIN 19266

Notes

- **All chemicals used for preparing this buffer solution are for analysis, ACS quality.**
- **The stability of prepared solution must be checked periodically by the user.**
- **Don't use the solution if it is contaminated accidentally.**

Panreac Química SLU
C/Garraf, 2
Polígono Pla de la Bruguera
E-08211 Castellar del Vallès
(Barcelona) Spain
Phone +34 937 489 400
info.es@itwreagents.com
www.itwreagents.com



QA Director
T. Llauradó

CODE: **273618** **BATCH:** 0002196890